

Instruction for Use







Flower Orthopedics Bone Screws Set Implants

Manufacturer:

FLOWER ORTHOPEDICS CORPORATION (FOC)

100 Witmer Road
Horsham, PA 19044

US: 1-877-778-8587
Intern.: +1 267-437-3063

Caution: Federal law restricts this device to sale by or on the order of a physician.	
	• Single use
	• Catalog number
	• Lot number
	• See instructions for use
	• Sterile
	• Use until

CONTENTS

The package contains one implant device of the Flower Orthopedics Bone Screws Set.

DESCRIPTION

The devices are supplied **sterile**. The devices are available in several sizes.

MATERIAL

The screws are made from titanium alloy (Ti6Al4V ISO 5832-3)

MRI SAFETY INFORMATION

The Flower Orthopedics Bone Screw Set has not been evaluated for safety and compatibility in the MR environment. It has not been tested for heating, migration, or image artifact in the MR environment. The safety of Flower Orthopedics Bone Screw Set in the MR environment is unknown. Scanning a patient who has this device may result in patient injury.

INTENDED USE

The Flower Orthopedics Flower Bone Screws set is intended to be used for the fixation of bone fractures, fusion of joints or bone reconstruction.

CONTRAINDICATIONS

Do not use the Flower Orthopedics implants in cases of:

- Inadequate bone quantity and/or bone quality
- Foreign body sensitivity to implant material
- Acute localized infections
- Patients with limited blood supply
- Patients with unstable physical and / or mental health conditions

ADVERSE REACTIONS

Adverse reactions may include but are not limited to:

- Clinical failure (i.e. pain or injury) due to bending, loosening, wear and tear, fracture of implant, loss of fixation, dislocation and/or migration
- Pain, discomfort, and/or abnormal sensations due to the presence of the implant
- Primary and/or secondary infections
- Allergic reactions to implant material
- Limb shortening due to compression of the fracture or bone resorption
- Necrosis of bone or decrease of bone density
- Injury to vessels, nerves and organs
- Hematoma and/or impaired wound healing; hemorrhage.

SAFETY PRECAUTIONS

- Each patient's record shall document the implant used (name, item number, lot number (if available)).
- Never re-use an implant. Although the implant may appear undamaged, previous stresses may have created non-visible damage that could result in implant failure. The manufacturer accepts no responsibility for a re-used implant.

HOW SUPPLIED/STORAGE:

The implants are individually packed in protective packaging that is labeled to its contents. All implants are supplied **sterile**.

- Always store the implants in the original protective packaging.
- Store the implants in a dry and dust-free place (standard hospital environment).

INSPECTION:

Before use, inspect the box carefully. Do not use when sterile barrier is visually damaged.

OPERATING INSTRUCTIONS

The Flower Orthopedics implants should be implanted only with the Flower Orthopedics instruments applicable for the respective sizes.

PRE-OPERATIVE

- Prior to use, thoroughly read the provided operation manual and become familiar with the surgical technique.
- Keep the instructions for use accessible to all staff.
- The operating surgeon must have a thorough understanding of both, the hands-on and conceptual aspects of the established operating techniques. Proper surgical performance of the implantation is the responsibility of the operating surgeon. The operating surgeon draws up an operation plan specifying and documenting the following:
 - Implant component(s) and their dimensions.
 - Determination of intra-operative orientation points.

The following conditions must be fulfilled prior to application:

- All required implant components are sterilized and readily available.
- All requisite sterile implantation instruments must be available and in working order.
- Highly aseptic operating conditions are present.

WARNING:

The use of implants for tasks other than those for which they are intended may result in damaged/ broken implants or patient injury.

- The operating surgeon and operating room team must be thoroughly familiar with the operating technique, as well as the range of implants and instruments to be applied. Complete information on these subjects must be readily available at the workplace.
- The operating surgeon must be especially trained in orthopedic surgery, biomechanical principles of the skeleton, and the relevant operating techniques.
- The operating procedure must be explained to the patient, and the patient's understanding of the following information must be documented:
 - The patient is aware of the risks associated with general surgery, orthopedic surgery, and with general anesthesia.
 - The patient has been informed about the advantages and disadvantages of the implant procedure and about possible alternative treatments.
 - The implant can fail due to excessive load, wear and tear or infection.
 - The service life of the implant is determined by body weight and physical activity. The implant must not be subjected to overload too early through extreme strain, work-related or athletic activities.
 - Corrective surgery may be necessary if the implant fails.
 - The patient must have his/her physician carry out follow-up examinations of the implants at regular intervals.

INTRA-OPERATIVE

- Prior to use, verify the integrity of the implant.
- Modification of the Flower Orthopedics Bone Screw Set is not allowed.
- Use the appropriate Flower G-Wire and Flower Drill (if necessary) to the appropriate length to prepare the path for the bone screws.

POST-OPERATIVE

- Reiterate preoperative instructions to the patient.
- During the post-operative phase, in addition to mobility and muscle training, it is of particular importance that the physician keeps the patient well informed about post-surgical behavioral requirements.
- Ensure that the patient is aware of physical activity restrictions and possible adverse reactions.

REVISION SURGERY / IMPLANT REMOVAL

Metallic implants can loosen, fracture, corrode, migrate, cause pain, or stress shield bone even after a fracture is healed, particularly in young, active patients. The surgeon must make the final decision on implant removal if either of these occurs. If there are not any of these complications, we recommend the permanent implantation of this titanium implants because of the risk of refracture and the possible complications of an additional operation.

Use of original products

Implants and instruments of the Flower Orthopedic Bone Screws Set are produced and designed to be used together.

WARNING

Please note that using a single use device (SUD) which comes into contact with human blood or tissue constitutes reprocessing & that reprocessing can only be performed by an authorized third-party reprocessor who has received 510(k) clearance for such.

WARRANTY

The manufacturer is not responsible for any complications arising from incorrect diagnosis, choice of incorrect implant, incorrect operating techniques, and limitations of treatment methods or inadequate asepsis.

All warranty rights are lost if repairs or modifications are carried out by an unauthorized service center. The manufacturer does not take responsibility for any effects on safety, reliability or performance of the product if the product is not used in conformity with the instructions for use. Technical alterations reserved.

FOR FURTHER INFORMATION

Please contact Flower Orthopedic Corporation if further information on this product is needed.



Flower Orthopedics Bone Screw Set - Instruments

Instruction for Use
Flower Orthopedics Bone Screw Set
Instruments

Manufacturer:

**FLOWER ORTHOPEDICS CORPORATION
(FOC)**

100 Witmer Road
Horsham, PA 19044

US: 1-877-778-8587
Intern.: +1 267-437-3063

Caution: Federal law restricts this device to sale by or on the order of a physician.

	• Single use
	• Catalog number
	• Lot number
	• See instructions for use
	• Sterile
	• Use until

CONTENTS

The package contains one or several surgical instrument(s) for use with the Flower Orthopedics Bone Screw Set.

DESCRIPTION

General instruments such as G-Wires, Drills, Drill Guides, Depth Gauges, and Screwdrivers.

MATERIAL

The instruments are made from medical stainless steel & plastic.

MRI SAFETY INFORMATION

The Flower Orthopedics Bone Screw Set has not been evaluated for safety and compatibility in the MR environment. It has not been tested for heating, migration, or image artifact in the MR environment. The safety of Flower Orthopedics Bone Screw Set in the MR environment is unknown. Scanning a patient who has this device may result in patient injury.

INTENDED USE

The Flower Orthopedics Flower Bone Screws set is intended to be used for the fixation of bone fractures, fusion of joints or bone reconstruction.

ADVERSE REACTIONS

Possible reactions may include but are not limited to:
Clinical failure due to inappropriate usage.
Necrosis due to thermal load (power driven tools).

SAFETY PRECAUTIONS

Prior to use, thoroughly read this instruction for use.
Keep the instructions for use accessible to all staff.
The use of a surgical instrument for tasks other than those for which they are intended may result in damaged or broken instruments or patient injury.

HOW SUPPLIED/STORAGE:

The instruments are packed in protective packaging that is labeled to its contents. All instruments are supplied **sterile**.

- Always store the instruments in the original protective packaging.
- Store the instruments in a dry and dust-free place (standard hospital environment).

INSPECTION:

Before use, inspect the instrument box carefully. Do not use when sterile barrier is visually damaged.

OPERATING INSTRUCTIONS

The Flower Orthopedics implants should be implanted only with the Flower Orthopedics instruments applicable for the respective sizes.

WARNING

Please note that using a single use device (SUD) which comes into contact with human blood or tissue constitutes reprocessing & that reprocessing can only be performed by an authorized third-party reprocessor.

WARRANTY

All warranty rights are lost if repairs or modifications are carried out by an unauthorized service center. The manufacturer does not take responsibility for any effects on safety, reliability or performance of the product if the product is not used in conformity with the instructions for use.
Technical alterations reserved.

FOR FURTHER INFORMATION

Please contact Flower Orthopedic Corporation or your authorized representative if further information on this product is needed.